

Manual bolus dose of insulin

S

A patient was given a manual bolus dose of insulin via a Braun syringe driver in error.

B

A

The drug library within the Braun syringe driver allowed a manual bolus of insulin to be given.

R

The information within the Drug Library for insulin on the Braun syringe drivers has been changed so that a manual bolus is now disabled.

Welcome to our safety learning bulletins, which aim to disseminate learning that has been shared from adverse incidents

We invite you to submit anonymous summaries of incidents that have occurred in your local units that have important lessons that we can all learn from to improve patient safety.

If you have an incident and learning that you would like to share please submit using the SBAR format.

We welcome any feedback regarding our shared learning process